Indirect ELISA kit for the detection of antibodies against APP serotypes 1 through 12 in swine serum, plasma and meat juice.

- Screening of 12 APP serotypes in a single test
 Serotyping may be performed on positive samples using the ID Screen[®] serotyping ELISAs
- Serum, plasma and meat juice protocols
- Fast, results in just 75 minutes
- The same straightforward protocol is used for all ID Screen® APP ELISAs, making it easier to use several tests at once

SPECIFICATIONS

Method	Indirect ELISA
Species	Swine
Specimens	Serum, plasma or meat juice
Coated antigen	APP antigens
Conjugate	Anti-swine IgG–HRP conjugate (concentrated 10X)

ORDERING INFORMATION

Product code	APPS-2P	APPS-10P
Kit format	2 plates	10 plates
Reactions	192	960
Plate format	12 x 8-well strips	12 x 8-well strips

REFERENCES

1. Pourquier, P., Lesceu, S. **Use of the ID Screen® APP indirect ELISA range for quick identification and serotyping of Actinobacillus pleuropneumoniae in domestic pigs**. Poster presented at the 6th International Symposium on Emerging and Re-emerging Pig Diseases, June 2011, Madrid, Spain.

ASSOCIATED PRODUCTS

APP 1-9-11 positive freeze-dried serum

Product code	MRI-APP-1-9-11
Format	0.25 ml vial (freeze-dried)
Description	Freeze-dried positive swine serum to be used as internal reference material for quality control.

APP 2 positive freeze-dried serum

Product code	MRI-APP-2
Format	0.25 ml vial (freeze-dried)
Description	Freeze-dried positive swine serum to be used as internal reference material for quality control.

APP 3-6-8 positive freeze-dried serum

Product code	MRI-APP-3-6-8
Format	0.25 ml vial (freeze-dried)
Description	Freeze-dried positive swine serum to be used as internal reference material for quality control.

APP 4-7 positive freeze-dried serum

Product code	MRI-APP-4-7
Format	0.25 ml vial (freeze-dried)
Description	Freeze-dried positive swine serum to be used as internal reference material for quality control.

APP 5 positive freeze-dried serum

Product code	MRI-APP-5
Format	0.25 ml vial (freeze-dried)
Description	Freeze-dried positive swine serum to be used as internal reference material for quality control.

APP 10 positive freeze-dried serum	
Product code	MRI-APP-10
Format	0.25 ml vial (freeze-dried)
Description	Freeze-dried positive swine serum to be used as internal reference material for quality control.

APP 12 positive freeze-dried serum

Product code	MRI-APP-12
Format	0.25 ml vial (freeze-dried)
Description	Freeze-dried positive swine serum to be used as internal
	reference material for quality control.

APP negative freeze-dried serum

Product code	MRINEG-APP
Format	0.5 ml vial (freeze-dried)
Description	Freeze-dried negative swine serum to be used as internal reference material for quality control.